

Manual Supplement

Manual Title: 323/324/325 Users
Part Number: 4045153
Print Date: May 2012
Revision/Date:

Supplement Issue: **2**
Issue Date: 9/12
Page Count: 1

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1, 62269

On page 3, add the following Warning:

- **Do not use a current measurement as an indication that a circuit is safe to touch. A voltage measurement is necessary to know if a circuit is hazardous.**

On page 10, under **Specifications**, *AC Current (Jaw)*, change Accuracy for the 324:

Change: Accuracy.....2.0 % ± 5 digits (45 Hz - 65 Hz)

2.5 % ± 5 digits (65 Hz - 400 Hz)

To: Accuracy.....1.5 % ± 5 digits (45 Hz to 400 Hz)

Add the following Note:

Note

Add 2 % for position sensitivity.

Change #2, 63287

On page 11, change:

From: *Frequency (325)*

To: *Frequency with Jaw (325)*